

VERCENTRA HA-30

Hematology Analyzer 3 Part Diff



Throughput

60 sampel/jam

Prinsip kerja

Impedansi untuk WBC, RBC, PLT
Metode kolorimetri non-Sianida
untuk HGB

Parameter

24 Parameter:

WBC, Lym#, Mid#, Gran#, Lym%,
Mid%, Gran%, NLR, PLR, RBC,
HGB, HCT, MCV, MCH, MCHC,
RDW-CV, RDW-SD, PLT, MPV,
PDW-CV, PDW-SD, PCT, P-LCC,
P-LCR

Histogram

3 Histogram (WBC, RBC, PLT)

Volume Sampel

Whole blood mode: 9µl

Capillary mode: 9µl

Prediluted mode: 20µl

Kalibrasi

Kalibrasi manual

Kalibrasi otomatis

Kalibrasi *fresh blood*

Kontrol Kualitas

3-Level QC

L-J QC

X-B QC

Lingkungan kerja

Rentang suhu lingkungan 10-30°C,

Kelembapan ≤70%, 100-240V,

50/60Hz, 160VA, 70kPa-106kPa

Reagen

Diluent, Lyse, Probe cleaner

Cetak

Auto print, Manual print

Komunikasi

Support Bi-directional LIS

Bahasa

Bahasa Indonesia

Bahasa Inggris

Input dan output

10,4 inch touch screen

Internal printer

Support external printer

Interface

USB port, Network port

Penyimpanan data

600.000 hasil pemeriksaan
dengan histogram

Dimensi

430mm x 280mm x 400mm
(PxLxT)

Berat

17,5kg

Kemenkes RI AKD 20205320441

Nilai TKDN

44,79%

Diproduksi oleh:

PT. VIRTUE DIAGNOSTICS INDONESIA

Didistribusikan oleh:

PT DARMA LABORA INDONESIA

Kawasan Industri Jababeka 7, Kav. B3B, Cikarang Utara, Kab. Bekasi, Jawa Barat 17534

☎ 021 38725029 ✉ customerservice.id@virtuedx.com 🌐 virtuedx.co.id

